ı	Application No.	Applicant(s)	
Notice of Allewshilling	10/733,491	YAMAGUCHI ET AL.	
Notice of Allowability	Examiner	Art Unit	
	L. G. Lauchman	2877	
The MAILING DATE of this communication appears on the cover sheet with the correspondence address All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.			
1. This communication is responsive to 2/03/2005.			
2.  The allowed claim(s) is/are 11-22.			
3.  The drawings filed on 10 December 2003 are accepted by the Examiner.			
4.			
Attachment(s)  1. Notice of References Cited (PTO-892)  2. Notice of Draftperson's Patent Drawing Review (PTO-948)  3. Information Disclosure Statements (PTO-1449 or PTO/SB/0 Paper No./Mail Date  4. Examiner's Comment Regarding Requirement for Deposit of Biological Material	5. Notice of Informal Page 1 Notice of Informal Page 1 No./Mail Dat Page 1 No./Mail Da	(PTO-413), e nent/Comment	·

## **DETAILED ACTION**

## Allowable Subject Matter

Claims 11-22 are allowed.

The following is an examiner's statement of reasons for allowance:

The Applicants canceled claims 1-10, according to the Amendment received 2/03/2005, added new claims 11-12, and presented arguments in favor of allowance of the existing claims.

As to Claim 1, the prior art of record taken along or in combination, fails to disclose or render obvious a single detector paired with the converging lens to detect the detecting light after the detecting light passes through a thermal lens generated in the fluid sample by the convergent irradiation of the exiting light, and synchronizing device which synchronizes an output signal of the detector with the modulation of the existing light by the modulator, in combination with the rest of the limitations of the claim. (see disclosure in the specification s at page 9, lines 19-34, page 10, lines 14-27, page 12, lines 9-11, page 16, lines 22-31, and page 17, lines 2-5)

The dependent Claims 2-22 are allowed by virtue of their dependence upon Claim 1.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Application/Control Number: 10/733,491

Art Unit: 2877

## Conclusion

Page 3

Any inquiry concerning this communication or earlier communications from the examiner should be directed to L. G. Lauchman whose telephone number is (571) 272-2418. The examiner's normal work schedule is 8:00am to 4:30pm (EST), Monday through Friday. If attempts to reach examiner by the telephone are unsuccessful, the examiner's supervisor Gregory J. Toatley, Jr. can be reached on (571) 272-2059, ext. 77.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Any inquiry of a general nature or relating to the status of this application should be directed to the TC receptionist whose telephone number is (571) 272-1562.

L. G. Lauchman Primary Examiner

Art Unit 2877

March 16, 2005